

Opal Viewer . Once in the Opal Viewer, go to: Settings >> Edit



In order to utilize the Zoom to Actual Size icon (), you must first calibrate your monitors within the

Under the Local Option tab:

OPAL Viewer Settings	
Local Options User Options Modality Options Annotations Options Chiropractic Options Overlay Config ToolBox Options	
Display Settings Monitors on System: 2 Monitors on System: 2 Monitor Settings Main Monitor Monitor Calibration Gray Width/mm 380 GrayOnly Height/mm 310 Color Color	Options Cache Path c:\opal\cache\viewer Keep In Cache For (Hours) 72 Check OnCall Every (Seconds) 150 Cache Uncompressed Images Hard Disk Swapping Occurs At (MB) 1228 Verify DM
Interface Options 1. Select each monitor to calibrate. 2. Using a ruler or other measurement device, measure the width and height of your monitor (<i>inside the display border</i>) in mm units. 3. Enter information, then hit OK	External Tools/Integration TraumaCADWeb TraumaCAD 3viseon Opal3D Title Description Command Arguments Inital Dir
	Add Update Remove
	OK Cancel Apply